



## ● MALDI Guided SpatialOMx uncovers proteomic profiles in tumor subpopulations of breast cancer

MALDI Guided SpatialOMx provides an excellent possibility to discover deep proteomics insights into heterogeneous tumor subpopulations by retaining the regiospecific information of an imaging technique.

### Abstract

The timsTOF fleX system bridges a current gap by providing MALDI Imaging and in-depth proteomics analysis in just one instrument. The instrument offers all benefits of a timsTOF Pro for time-efficient and sensitive proteomics, combined with a high-resolution

MALDI source and stage. Using PASEF technology, it is possible to retrieve high protein ID rates with small sample amounts. Here we present the new SpatialOMx workflow to identify distinct proteomic profiles for different tumor subpopulations in breast cancer as an example for this powerful approach.

### Introduction

Since tissue and disease are correlated, SpatialOMx provides the unique opportunity to combine regiospecific information from MALDI Imaging with deep proteomic coverage for biomarker discovery and molecular characterization. MALDI Imaging has

*Keywords:*  
SpatialOMx, MALDI  
Imaging, Proteomics,  
PASEF

the advantage of providing spatial and molecular information that can be used to detect Region of Interests (ROIs) in tissue sections. As a discovery tool, MALDI Imaging lacks the possibility of in-depth molecular characterization of classical LC-MS/MS based OMICs methods. Classical OMICs strategies, however, do not retain the spatial information.

Here, we present SpatialOMx for microproteomic characterization of tumor subpopulations in breast cancer. In this efficient workflow, unsupervised segmentation of MALDI Imaging data with SCiLS Lab is used to define ROIs. Image processing strategies provide boundary information of segments of tumor subpopulations for laser capture microdissection (LMD). Protein extraction and tryptic digestion of small microdissected material is followed by proteomic analysis on the timsTOF fleX using PASEF (Parallel Accumulation Serial Fragmentation). Analysis of proteomics data provides a comprehensive mechanistic understanding of cell-type specific biological processes *in-situ* to complement the workflow.

## Methods

### MALDI Imaging analysis, staining and clustering analysis

Fresh breast cancer tissue was collected by the Tissue Biobank of the University of Liège, directly frozen in liquid nitrogen, and then stored at  $-80^{\circ}\text{C}$ . The standardized protocol was approved by the Ethics Committee of the University Hospital Center of Liège. Informed consent was obtained from the participant included in this study.

Fresh-frozen breast tumor sections were cut at  $12\ \mu\text{m}$  thickness and mounted on PEN (polyethylene naphthalate) membrane slides (Leica Microsystems). These slides are

compatible with MALDI Imaging and with LMD. After drying, the slides were coated with norharmane matrix (7 mg/mL in  $\text{CHCl}_3:\text{MeOH}$ , 2:1, v:v) using a TM-sprayer (HTX Technologies) for lipid analyses. 12 matrix layers were applied with a 30 s break between each layer at a flow rate of 0.120 mL/min, a velocity of 1200 mm/min and a nozzle temperature of  $30^{\circ}\text{C}$ . The slides were scanned with a slide scanner (TissueScout, Bruker) to obtain a grayscale reference image with teach marks. Lipid MALDI Imaging analysis was performed on a timsTOF fleX instrument over the entire section with  $50\ \mu\text{m}$  spatial resolution in the mass range 300-1600  $m/z$ .

After MALDI Imaging analysis, the matrix was removed, and the section was stained with hematoxylin and eosin, scanned with a digital slide scanner for pathological evaluation, and coregistered to the MALDI Imaging data using the previously applied teach marks. MALDI Imaging data were imported into SCiLS Lab 2019c. Spectra from the pathological annotation of the tumor region were analyzed by unsupervised clustering using the *k*-means algorithm to define

three tumor subpopulation regions of interest (ROIs).

### Image processing of cluster results and LMD

As segmentation by *k*-means provides fine structures that are not processable by LMD, the segmentation image data had to be smoothed. Therefore, segmentation data of the tumor subpopulation areas were imported to MatLab R2018a for smoothing, removal of small objects, holes filling, and boundary detection. Hereafter, the different segmentation images were up-scaled to the resolution of the optical image. This upscaled segmentation information was then transferred to a consecutive unstained, unmeasured section located on a separate PEN slide. This was achieved through coregistration of both sections using prominent morphological features of the sample. The untouched section was also surrounded by teach marks which are visible in the LMD and can therefore function as reference points for the recalculation of the boundary coordinates to the coordinate system of the LMD. Areas containing approximately 2000 cells were




Figure 1: Image of the H&E stained section with tumor area marked in blue by a pathologist (left side). Zoom on the tumor area after segmentation analysis using *k*-means clustering, which produces three segments representing tumor subpopulations (right side).




Figure 2: Image processing pipeline to produce ROI boundary information as  $x,y$ -coordinates for laser capture microdissection from segmentation raw data. The workflow includes smoothing, removal of small objects, holes filling, boundary detection and up-scaling to the optical image.

dissected from each segmented tumor subpopulation with a Leica LMD7000 and directly transferred to reaction tubes for microextraction and tryptic digestion.

#### Sample preparation for proteomics

Samples for protein identification by PASEF on the timsTOF fleX were processed as described [2]. Briefly, extraction solution was added, and samples were heated at 99°C for 1 h. The pH was adjusted to 7.4, samples were reduced and alkylated and digested overnight at 37°C with trypsin. After digestion, samples were dried and desalted using ZipTip purification.

Proteomics data acquisition using PASEF technology on the timsTOF fleX

Microextracted peptide samples were run on a Bruker nanoElute system using a one column setup equipped with an Aurora 25 cm x 75 μm 1.7 μm C18 column (IonOpticks, Parkville, Australia). 0.1% formic acid,

2% acetonitrile in water was used as eluent A, and 0.1% formic acid in acetonitrile as eluent B with a 400 nL/min gradient at 50°C oven temperature. The Bruker CaptiveSpray nano source was used as inlet into the timsTOF fleX system using 3.0 L/min dry gas at 180°C and 1.6 kV capillary voltage. Data were acquired in the mass range of 100–1700  $m/z$  and the mobility range was set to 0.6–1.6  $1/k_0$ . For MS/MS fragment spectra acquisition, 10 PASEF scans at 100 ms were run, from which each contains around 12 MS/MS events resulting in a > 100 Hz acquisition rate.

#### Analysis of proteomics data

PEAKS X (Bioinformatics Solutions Inc, Waterloo, Canada) was used for protein identification. Database searches were carried out against a SwissProt human database using a parent mass error tolerance of 15 ppm and a fragment mass tolerance of 50 mDa. Trypsin was used as proteolytic enzyme and maximally 2 missed cleavages were allowed. Carbamidomethylation was set as

fixed modification, acetylation at the protein N-terminus, deamination (NQ) and oxidation (M) were allowed as variable modifications. For biological process characterization of the different tumor subpopulations, the gene ontology tool PANTHER V.13.1 was used.

## Results

### MALDI Imaging defines tumor subpopulation ROIs

Lipid imaging on the timsTOF fleX from tumor sections was used to define ROIs by segmenting intratumor heterogeneity for laser microdissection (LMD) based microproteomics. The workflow is supported by the compatibility of PEN (polyethylene naphthalate) membrane slides with MALDI Imaging on the timsTOF fleX. These slides also support histological staining and are compatible with LMD. Lipid MALDI Imaging has the advantage that the same tissue section can be used for imaging, staining and LMD based microproteomics. Moreover, discernible lipid

profiles are the basis for segmenting intratumor heterogeneity.

The breast tumor section was stained with hematoxylin and eosin (H&E) after MALDI Imaging data acquisition and tumor areas were annotated by a pathologist. MALDI Imaging data with 50 µm spatial resolution were imported into SCiLS Lab 2019c and the annotated tumor areas analysed by unsupervised clustering using the SCiLS Lab segmentation algorithm. K-means segmentation highlighted three subpopulations in the annotated tumor region (Figure 1).

ROI boundary coordinates are the basis for precise LMD

Fine structure segmentation ROIs require image processing to allow viability of LMD. Currently, the segmentation image raw data contains small objects or edged segment boundaries that prevent direct usage of the coordinates for LMD. After image processing, boundaries of the tumor subpopulation areas were transferred to a coregistered optical image of a consecutive section and referenced to teach marks that were scanned together with the section (Figure 2). This allows co-registration

of the MALDI Imaging segment boundaries to the LMD for precise excision of the defined areas.

Proteomics of microextracted MALDI Imaging segments by PASEF technology implemented in timsTOF fleX

The timsTOF fleX system supports PASEF (Parallel Accumulation Serial Fragmentation) technology for time efficient and sensitive X-omics analysis. PASEF takes advantage of the dual TIMS device which separates incoming ion packages from one scan according to their ion mobility and elutes these ions in a time focused manner. While ions from the continuous ESI process are separated and eluted in the second TIMS device, ions are accumulated in parallel in the first part of the dual TIMS cartridge allowing for a 100% duty cycle. Alternating precursor and fragment ion scans complete assignment of fragment ion spectra to their respective precursor spectra. PASEF allows for scan rates of more than 100 Hz without losing sensitivity or resolution and is the optimal system for small sample amounts and therefore for SpatialOMx.

Tryptic digests of the extracted proteins from the excised MALDI Imaging tumor subpopulation areas were run on the timsTOF fleX using 100 min gradients on a Bruker nanoElute LC. The excised ROIs consisted of approximately 2000 cells per sample. On average, about 3500 protein IDs were obtained from the three microextractions to provide a deep insight into the proteome from just 160 ng (estimated) of sample. The molecular properties of the different tumor subpopulations were characterized and represented as Venn diagrams (Figure 3).

Interestingly, the proteomic profiles of the tumor subpopulation differed largely. The gene ontology terms “Biological regulation” and “Developmental processes” were underrepresented in subpopulations 1 and 2, respectively, while “cellular component organization” was over-represented in tumor subpopulation 3.

In summary, SpatialOMx provides guidance to specific regions of interest for microextraction and in-depth proteomics analysis to characterize regiospecific molecular changes in detail.




Figure 3: Proteomics of tumor subpopulations and biological process characterization. Proteins from microdissected tissue (approx. 160 ng) were extracted, digested with trypsin and peptide extracts were run on the timsTOF fleX using PASEF. Number of protein IDs per tumor subpopulation segment (A) and biological process characterization per segment as revealed by PANTHER (B).

## Conclusion

- SpatialOMx is a new workflow for *in-situ* characterization of tissue subtypes.
- MALDI Imaging provides guidance for laser-capture microdissection.
- SpatialOMX, combining MALDI Imaging and LC-MS/MS, can be performed in just one instrument: timsTOF fleX.
- timsTOF fleX allows for in-depth proteomic analysis of small microdissected tissue pieces using PASEF technology.



## Learn More

You are looking for further Information?  
Check out the link or scan the QR code for more details.

[www.bruker.com/timstofflex](http://www.bruker.com/timstofflex)



### References

- [1] Meier F et al., (2015), J. Proteome Res., **14**, 5378-5387
- [2] Longuespée R et al., (2016), Methods, **104**, 154-162
- [3] Dewez F et al., (2019), Anal. Bioanal. Chem, **411**, 5647-5633

For Research Use Only. Not for Use in Clinical Diagnostic Procedures.

● **Bruker Daltonik GmbH**

Bremen · Germany  
Phone +49 (0)421-2205-0

**Bruker Scientific LLC**

Billerica, MA · USA  
Phone +1 (978) 663-3660

[ms.sales.bdal@bruker.com](mailto:ms.sales.bdal@bruker.com) – [www.bruker.com](http://www.bruker.com)