

Application Note

Comprehending COVID-19: Reversed-Phase Liquid Chromatography (RPLC) of Intact SARS-CoV-2 Spike Protein

Jennifer M. Nguyen, Matthew A. Lauber

Waters Corporation



Save 15% off on Columns, Consumables and Spare Parts on Waters.com. Use code APP15 and start saving today. Terms & Conditions may apply.

Need Help? To learn more about how Waters can help you in your efforts against COVID-19, please contact the [COVID-19 Innovation Response Team](#)

This is an Application Brief and does not contain a detailed Experimental section.

Abstract

The global COVID-19 pandemic has resulted in extensive efforts to develop vaccines for the novel coronavirus. Identifying vaccine targets relies on robust analytical methods to understand SARS-CoV-2 structural biology. This study focuses on reversed-phase liquid chromatographic analysis of the intact SARS-CoV-2 spike protein, which has emerged as a potential target for vaccine development due to its role in viral pathogenesis.^{1,2} This work demonstrates that using difluoroacetic acid (DFA) as a mobile phase modifier in place of formic acid (FA) results in increased chromatographic resolution during intact protein analysis. Furthermore, the results suggest that pairing this approach with N- and O-glycosidase treatments may enable more detailed intact protein MS investigations.

Benefits

Using DFA instead of FA as the mobile phase modifier achieves:

- Higher resolution of less abundant proteoforms
- Three-fold increase in gradient peak capacity

Introduction

The SARS-CoV-2 spike protein, which facilitates host cell infection, has become a subject of detailed study due to its potential as a COVID-19 vaccine target. Proper characterization of this novel coronavirus protein relies on robust identity and purity tests. While extensive characterization work is underway to study the SARS-CoV-2 spike protein's glycans and glycopeptides, intact protein analysis using reversed-phase liquid chromatography (RPLC), either with or without the combined use of endoglycosidases, may offer unique analytical insights.^{3,4}




Figure 1. The SARS-CoV-2 spike protein (gray) with glycans modeled on its surface. Lorenzo Casalino, Zied Gaieb, and Rommie Amaro, UC San Diego.

To aid this effort, Waters shares the following method:

- A comparison of an intact RPLC profile using mobile phases modified with either difluoroacetic acid (DFA) or formic acid (FA). DFA is shown to enhance resolving power while maintaining MS-compatibility.

Experimental

The following experimental conditions were used for RPLC-FLR-MS intact protein analysis of the SARS-CoV-2 spike protein.

LC Conditions

LC system:	ACQUITY UPLC I-Class
Detection:	FLR (280 nm emission, 320 nm excitation)
Vials:	QuanRecovery vials
Column(s):	BioResolve RP mAb Polyphenyl, 2.7 μm , 450 \AA , 2.1 x 50 mm
Column temp.:	80 °C
Sample temp.:	8 °C
Injection volume:	1 μL
Flow rate:	0.2 mL/min
Mobile phase A:	0.1% IonHance DFA or FA in water
Mobile phase B:	0.1% IonHance DFA or FA in acetonitrile
Gradient:	15–55% Mobile phase B in 20 minutes

MS Conditions

MS system: Vion IMS QToF Mass Spectrometer

Ionization mode: ESI+

Acquisition range: 1500–4000 *m/z*

Capillary voltage: 2.25 kV

Collision energy: 6 V

Cone voltage: 140 V

Results and Discussion




Figure 2. Comparison of FA and DFA Intact Protein Fluorescence Signal.




Figure 3. Comparison of FA and DFA Intact Protein Total Ion Chromatograms.

Employing DFA as a mobile phase modifier resulted in a comparatively higher resolution chromatogram. Compared to using FA as a mobile phase modifier, gradient peak capacity increased by over three-fold while the less abundant proteoforms were better resolved. Pairing this chromatographic approach with N- and O-glycosidase treatments may enable more detailed MS investigations at the intact protein level of analysis.

Conclusion

Because the SARS-CoV-2 spike protein is implicated in viral pathogenesis, it has become a target for vaccine development. Efficient therapeutic development relies on a solid structural and functional understanding of the SARS-CoV-2 spike protein target. Intact protein analysis using RPLC can be used to refine our understanding of the SARS-CoV-2 spike protein and thus help to identify and develop promising new COVID-19 therapies. This work demonstrates that the use of DFA instead of FA as mobile phase modifier enhances method resolving power while maintaining MS-compatibility.

References

1. Pinto, D. *et al.* Structural and Functional Analysis of a Potent Sarbecovirus Neutralizing Antibody. *bioRxiv* 2020.04.07.023903 (2020). doi: <https://doi.org/10.1101/2020.04.07.023903>
2. Stawiski, E.W. *et al.* Human ACE2 Receptor Polymorphisms Predict SARS-CoV-2 Susceptibility. *bioRxiv* 2020.04.07.024752 (2020). doi: <https://doi.org/10.1101/2020.04.07.024752>
3. Liu, X. and Lauber, M. Comprehending COVID-19: Rapid and Sensitive Characterization of N-Glycans from SARS-CoV-2 Spike Protein. Waters Application Highlight [720006914](#).
4. Novokmet, Mislav *et al.* Understanding Glycans in COVID-19 Drug Design. <https://www.genengnews.com/insights/understanding-glycans-in-covid-19-drug-design/>

Featured Products

[ACQUITY UPLC I-Class PLUS System](#)

[ACQUITY UPLC FLR Detector](#)

[Vion IMS QToF Ion Mobility Quadrupole Time-of-flight Mass Spectrometry](#)

720006907, Revised December 2020

©2020 Waters Corporation. All Rights Reserved.