


Application Note

## Analysis of Vitamins A and E in Serum by UPLC-QDa for Clinical Research

---

Karl Lo, Fionn Quinlan, Mark Ritchie

Waters Corporation



This is an Application Brief and does not contain a detailed Experimental section.

---

## Abstract

Here we describe a clinical research method for the simultaneous analysis of vitamins A and E in serum by single quadrupole mass detector, the ACQUITY QDa. The extraction method is a modification of a previous LC-MS/MS method, using HLB Prime for sample preparation. The combination of a chromatographic method that uses an ACQUITY UPLC HSS PFP Column and an ACQUITY QDa testing under SIR mode provides an accurate and sensitive method for the analysis of the vitamins A and E in serum. The injection cycle is just 5 minutes and the sensitivities could achieve 100 ppb and 1 ppm for vitamins A and E in the serum, respectively.

## Benefits

- An accurate and sensitive method for the analysis of the vitamins A and E in serum using a single quadrupole mass detector, the ACQUITY QDa and an ACQUITY UPLC H-Class System
- Simple sample preparation method compared to LLE

---

## Introduction

Historically, the majority of vitamins A and E analysis is performed by HPLC with UV detection. Even when the analysis of vitamin A and E involves LC-MS/MS, the extraction methods used such as liquid-liquid extraction (LLE) suffer from very long extraction times, with high solvent consumption and lengthy analysis time. Previously, we have shown a clinical research method on LC-MS/MS that requires just 100  $\mu$ L sample volume and provides a shortened analysis time.<sup>1</sup> The removal of phospholipid by Oasis HLB PRiME (SPE) is a key factor in this improvement of performance. In this technical brief, we will demonstrate how LC-MS could provide an accurate and sensitive analysis of vitamin A and E from the serum.

---

## Experimental

Isotopically labelled internal standards were added to the 100  $\mu$ L serum and then a protein precipitation procedure was performed by the addition of ethanol. After centrifugation, the supernatant was diluted by

ethanol/water (5:3) to give a final ethanol/water ratio of 2:3. The resulting solution of 650  $\mu$ L was applied to Oasis HLB elution plates for sample clean up. The sample was cleaned with 20% ACN (aq) and then eluted in 100% ACN. The eluent was further diluted with water to 200  $\mu$ L (Figure 1). Taking the advantage of using HLB PRiME, no activation and evaporation step was required. The 20  $\mu$ L resulting extract was injected to LC-MS for the analysis.




Figure 1. Chromatograms of 100 ppb of vitamin A and 1 ppm of vitamin E in serum.

Using an ACQUITY UPLC H-Class PLUS System with an ACQUITY HSS PFP Column (1.8  $\mu$ m, 2.1 x 50 mm, P/N:186005965 <<https://www.waters.com/nextgen/us/en/shop/columns/186005965-acquity-uplc-hss-pfp-fluoro-phenyl-column-100a-18--m-21-mm-x-50-.html>> ), a gradient elution of 65% mobile phase A (MP A) to 100% MP A was run to achieve the separation. Detection was performed by ACQUITY QDa under ESI positive and SIR mode for the corresponding protonated adducts. The injection cycle was 5 minutes.

## LC Conditions

LC system:	ACQUITY UPLC H-Class Parameter
Vials:	96-well collection plate containing 1 mL inserts (P/N: 186000855)

Column ACQUITY UPLC HSS PFP Column 2.1 x 50 mm,  
1.8 µm (P/N: 186005965)

Column temp.: 40 °C

Sample temp.: 10 °C

Injection volume: 20 µL

Flow rate: 0.40 mL/min

Mobile phase A: Water + 2 mM NH<sub>4</sub>Ac + 0.1% formic acid

Mobile phase B: MeOH + 2 mM NH<sub>4</sub>Ac + 0.1% formic acid

## Gradient

Time (min)	Flow (mL/min)	%A	%B	Curve
Initial	0.40	35	65	Initial
2.00	0.40	2	98	6
3.30	0.40	35	98	6
3.50	0.40	35	65	6
5.00	0.40	35	65	6

## MS Conditions

MS system: ACQUITY QDa

Ionization mode: ESI+

Acquisition mode	SIR
Capillary voltage:	0.8 kV
Cone voltage:	See channel details

	SIR, <i>m/z</i>	Cone (V)	Time
Vitamin A	269.15	30	0 to 2.5
D6 Vitamin A acid (VA IS)	307.2	10	
Vitamin E	431.40	5	2.5 to 4
D6 Vitamin E (VE IS)	437.4	5	

## Data Management

LC-MS software:	MassLynx v4.2
-----------------	---------------


---

## Results and Discussion

Figure 2 shows the chromatograms of 100 ppb and 1 ppm of vitamin A and E in the stripped serum (MSG2000). The analytical sensitivity investigations demonstrate that quantitation at 100 ppb of vitamin A and 1 ppm of vitamin E are achievable (%RSD <20, Bias <15%, S/N >10). The method was shown to be linear across the range of 100 ppb to 2000 pp for vitamin A and 1 ppm to 20 ppm for vitamin E. Correlation coefficients ( $r^2$ ) >0.993 achieved for both compounds across 10 separate occasions. Example of calibration curves are shown in Figure 2.

Correlation coefficient:  $r = 0.998153$ ,  $r^2 = 0.99609$   
 Calibration curve:  $0.00187808 * x + -0.0252769$   
 Response type: Internal Std (Ref 3), Area \* (IS Conc. / IS Area)  
 Curve type: Linear, Origin: Exclude, Weighting: 1/x, Axis trans: none

Vitamin A



Correlation coefficient:  $r = 0.999497$ ,  $r^2 = 0.998995$   
 Calibration curve:  $0.432241 * x + -0.883944$   
 Response type: Internal Std (Ref 4), Area \* (IS Conc. / IS Area)  
 Curve type: Linear, Origin: Exclude, Weighting: 1/x, Axis trans: none

Vitamin E




Figure 2. Examples of calibration curves of vitamins A and E.

Precision was assessed by extracting and analysing 5 replicates from low to high concentrations in stripped serum over 5 days ( $n=25$ ). Repeatability and total precision were  $\leq 8\%$  CV at all concentration levels tested for vitamin A and E and they were summarised in Table 1. All calculated % recoveries were within  $\pm 15\%$  for typical endogenous interferences tested when comparing test and control samples.

Compound	Total QC precision (RSD)			QC repeatability (RSD)		
	QC1 (low)	QC2 (mid)	QC3 (high)	QC1 (low)	QC2 (mid)	QC3 (high)
Vitamin A	7.2	7.6	7.3	3.4	3.6	4.4
Vitamin E	7.4	4.6	3.4	5.4	2.2	2.0

Matrix factor results were within  $\pm 15\%$  between test samples from six individuals and control samples compensated for using the internal standard (Table 2). Matrix suppression was observed for vitamin E (mean=0.653 and 0.730), but it was compensated by the internal standard (mean= 0.977 and 0.939).

Compound	Spiked conc.	Matrix factor – absolute peak area (range)	Matrix factor – adjusted conc. (range)
Vitamin A	300 ppb	0.932 (0.885 – 0.982)	0.945 (0.913 – 0.965)
	1600 ppb	0.919(0.875 – 0.971)	0.948 (0.916 – 0.992)
Vitamin E	3 ppm	0.653 (0.607 – 0.723)	0.977 (0.917 – 1.045)
	16 ppm	0.730 (0.6510 – 0.802)	0.939 (0.925 – 0.969)

---

## Conclusion

We explored the application of a UPLC-Single Quadrupole MS System, the ACQUITY UPLC-ACQUITY QDa, on a clinical research analysis of fat-soluble vitamins. The results show that the method is accurate and analytically sensitive even in a complex matrix. The sample preparation is relatively simple compared to LLE and the run time is significant decreased compare to traditional HPLC-UV methods.

---

## References

1. Simultaneous Analysis of Vitamin A and E in Serum by UPLC-MS/MS for Clinical Research. Waters Corporation, UK, 2019. Waters Application Note [720006642EN](https://www.waters.com/nextgen/us/en/library/application-notes/2019/simultaneous-analysis-vitamins-a-e-serum-uplc-ms-ms-clinical-research.html) <<https://www.waters.com/nextgen/us/en/library/application-notes/2019/simultaneous-analysis-vitamins-a-e-serum-uplc-ms-ms-clinical-research.html>> .

---

## Featured Products

ACQUITY UPLC H-Class PLUS System <<https://www.waters.com/10138533>>

ACQUITY QDa Mass Detector <<https://www.waters.com/134761404>>

MassLynx MS Software <<https://www.waters.com/513662>>

TargetLynx <<https://www.waters.com/513791>>

720007164, February 2021

© 2021 Waters Corporation. All Rights Reserved.