

# Analysis of Bovine Serum Albumin (BSA) Protein Digest on a Thermo Scientific Accucore 150-C18 150 Å Pore Diameter HPLC Column

Valeria Barattini, Joanna Freeke, Thermo Fisher Scientific, Runcorn, Cheshire, UK

## Key Words

Accucore C18, bottom-up proteomics, proteins, peptides, fused core, superficially porous, 150 Å, bovine serum albumin (BSA)


## Abstract

This application note demonstrates the analysis of trypsin-digested BSA using a Thermo Scientific Accucore 150-C18 HPLC column (150 Å pore diameter). The analysis is carried out using an acidified water:acetonitrile gradient over 40 minutes, with pressures compatible with conventional HPLC instrumentation. The data shows high peak capacity with excellent reproducibility.

## Introduction

Accucore™ HPLC columns use Core Enhanced Technology™ to facilitate fast and high efficiency separations. The 2.6 µm diameter particles have a solid core and a porous outer layer. The optimized phase bonding creates a series of high coverage, robust phases. The tightly controlled 2.6 µm diameter of Accucore particles results in much lower backpressures than typically seen with sub 2 µm materials. Accucore 150-C18 columns have been further optimized for the analysis of biomolecules and protein digests by bonding C18 ligands onto the porous outer layer, which has a pore size of 150 Å.

The sample preparation for a standard bottom-up proteomics experiment involves enzymatic digestion of proteins or even complex mixtures of proteins, such as a whole cell lysate. Breaking down an already complex sample mixture into a large number of peptide fragments of varying size and hydrophobicity presents an analytical challenge. It is crucial that the chromatographic separation is highly efficient with high peak capacity in order to achieve accurate and reliable identification of all the fragments present in the sample. The 150 Å pore diameter enables larger peptides to penetrate more efficiently into the stationary phase particle. This leads to improved interaction with the stationary phase surface and as a result increased resolution compared to a narrow pore diameter particle. The combination of Core Enhanced Technology with wide-pore silica, results in very high efficiency separations using conventional HPLC systems.



Developments in MS technology have enabled detection and analysis of very low abundance species within complex peptide mixtures. However, greater confidence in peptide identification can be achieved through the use of higher resolving power stationary phases.

In this application note we demonstrate the excellent performance of an Accucore 150-C18 HPLC column for the chromatographic separation of trypsin-digested BSA.

## Experimental Details

Consumables	Part Number
Fisher Scientific HPLC grade water	W/0106/17
Fisher Scientific HPLC grade acetonitrile	A/0626/17
Fisher Scientific Optima formic acid	A117-50
Proteabio BSA Digest Standard (lyophilized, 500 pmol)	PS-204-3
Liquid handling hardware: FinnPipette Kit 1	4700870
Vials and closures	MSCERT4000-34W

Separation Conditions	Part Number	
Instrument:	Thermo Scientific Accela HPLC System	
Column:	Accucore 150-C18, 2.6 $\mu\text{m}$ , 100 x 2.1 mm	16126-102130
Mobile phase A:	0.1 % formic acid in water	
Mobile phase B:	0.1 % formic acid in acetonitrile	
Gradient:	Time (min)	%B
	0	5
	5	5
	45	50
	46	95
	50	95
	51	5
	60	5
Flow rate:	0.2 mL/min	
Backpressure at starting conditions:	170 bar	
Run time:	60 minutes	
Column temperature:	40 °C	
Injection details:	3 $\mu\text{L}$ 5 pmol/ $\mu\text{L}$ solution	
Injection wash solvent:	water + 0.1% formic acid	

## MS Conditions

Thermo Scientific LCQ-Deca XP ion-trap mass spectrometer	
<b>Ionization</b>	ESI positive mode
<b>Sheath Gas Flow Rate</b>	60
<b>Aux Gas Flow Rate</b>	20
<b>Capillary Temperature</b>	300 °C
<b>Voltage</b>	4.5 kV
<b>Cap V</b>	30 V
<b>Mass Range</b>	300-1500 m/z

## Solutions

Standard preparation:	Digested and lyophilized BSA (500 pmol) was reconstituted in 100 $\mu\text{L}$ of water + 0.1% formic acid to give a 5 pmol/ $\mu\text{L}$ solution. The mixture was sonicated for 20s to ensure full solubilisation with minimal peptide aggregation.
-----------------------	--

## Data Processing

Software:	Thermo Scientific Xcalibur 2.0 SR2
-----------	------------------------------------

## Results

Analysis of BSA digest in the conditions described above leads to elution of tryptic peptides within 30 minutes.

The base peak chromatogram obtained for this separation is shown in Figure 1.




Figure 1: Base Peak chromatogram of digested BSA on Accucore 150-C18, 100 x 2.1 mm HPLC column

Retention time reproducibility was assessed by running triplicate analyses (Figure 2, base peak chromatograms). The retention times of a set of eight peptides from early, middle and late eluting regions were monitored for statistical analysis. In all eight cases excellent retention time reproducibility was observed, with %RSD values below 0.4 % (Table 1).




Figure 2: Base peak chromatogram of triplicate analyses of BSA digests

Peptide m/z	Retention time Average (min)	%RSD
395.33	14.97	0.37
523.73	15.90	0.19
656.13	16.29	0.29
547.73	19.57	0.19
628.00	20.23	0.32
708.87	23.15	0.15
534.67	23.66	0.20
625.73	25.01	0.12

Table 1: Retention time average and %RSD values for a set of eight peptides

### Resolution Power – Peak Capacity

For many analysts running complex digested samples, the peak capacity (i.e. the number of peaks that can be resolved within a gradient run) is a crucial parameter, which ultimately influences the number of peptide identification hits.

Peak capacity is conventionally calculated as follows [1]:

$$n_c = 1 + \left( \frac{t_g}{W} \right)$$

Where  $n_c$  = Peak capacity

$t_g$  = gradient time

$W$  = average peak width 10% height

The peak widths at 10% height were measured for peptides listed in Table 1; in all cases these were found to be excellent, with an average width of 10.8 s (Figure 3). A peak capacity calculation based on this data set resulted in an average value of 225, indicating the high resolution power of Accucore 150-C18 HPLC column.





Figure 3: EIC for two peptides in the monitoring set of eight. In both cases excellent peak widths are found

### Conclusion

- Accucore 150-C18 HPLC columns shows excellent separation of BSA digest fragments
- High peak capacity observed for a complex peptide digest
- Outstanding run-to-run reproducibility
- Backpressure compatible with conventional HPLC systems

### References

[1] Wang X., Peak Capacity Optimization of Peptide Separations in Reversed-Phase Gradient Elution Chromatography (2006), 78 (10), 3406-3416

[thermoscientific.com/chromatography](http://thermoscientific.com/chromatography)

© 2012 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific Inc. and its subsidiaries. Specifications, terms and pricing are subject to change. Not all products are available in all countries. Please consult your local sales representative for details.

**USA and Canada** +1 800 332 3331  
**France** +33 (0)1 60 92 48 34  
**Germany** +49 (0) 2423 9431 20 or 21  
**United Kingdom** +44 (0)1928 534110  
**Japan** +81 3 5826 1615

**China** +86 21 68654588 +86 10 84193588  
 +86 20 83145199 800 810 5118  
**India** +91 22 6742 9494  
**Australia** 1 300 735 292 (free call domestic)  
**New Zealand** 0800 933 966 (free call domestic)  
**All Other Enquiries** +44 (0) 1928 534 050

**Technical Support**  
**North America** +1 800 332 3331  
**Outside North America** +44 (0) 1928 534 440

**Thermo**  
 SCIENTIFIC

Part of Thermo Fisher Scientific