


Protein Mixture by Reversed-Phase HPLC on a Thermo Scientific™ Acclaim™ 300 C18 Column



20407

Many small soluble proteins are amenable to reversed-phase HPLC analysis. The wide 300- \AA pores of the Acclaim 300 column allow small-to-medium-sized proteins to be separated without band broadening due to size exclusion effects.