Biotage[®] Extrahera[™] LV-200 GLP Software Checklist

The following document will ensure that the GLP software is correctly installed and that all necessary tasks have been performed. For network configuration and other settings refer to the User Guide.

Please use the checklist below to track progress of the software installation and fill out this form electronically. A printed copy should be signed by both the Field Service Engineer and customer and provided to the customer after completion.



Installation Checklist

Software installtion

\checkmark	Tasks	Notes
	Install the GLP Software	
	Install the GLP Licence	Ensure that the system is installed with the latest version of Software.

Software Configuration, Verification and Functional Testing

\checkmark	Tasks	Notes
	Create an account with a system administrator access level and activate the account. Record the details.	Make sure that the system administrator is present when setting the user name, user password and email (optional).
		Password: The user's password for logging in to the system and the remote viewer (if enabled). The password must contain between 8 and 40 characters, at least one upper case and one lower case character, and at least one number.
	Set the default expiry date to 120 days.	After 120 days customer will recieve a notification to reset the password.
	Ensure that the system administrator can log in with the correct credentials.	Log out and log back in.
	Verify that the network port (LAN) is functional.	Check the functionality by connecting to your laptop and ping the IP address.



Record Section

System Information						
Extrahera Serial#						
Application SW Version						
User First Name						
User Last Name						
Email (optional)						
Password						

Notes

This checklist confirms installation has been provided as outlined in this document.

Company Name and Location		
Customer Name	Date	
Field Service Engineer	Date	

EUROPE	NORTH & LATIN AMERICA	JAPAN	KOREA	Distributors in other regions
Main Office: +46 18 565900	Main Office: +1 704 654 4900	Tel: +81 3 5627 3123	Tel: +82 31 706 8500	are listed on www.biotage.com
Fax: +46 18 591922	Toll Free: +1 800 446 4752	Fax: +81 3 5627 3121	Fax: +82 31 706 8510	
Order Tel: +46 18 565710	Fax: +1 704 654 4917	jp_order@biotage.com	korea_info@biotage.com	
Order Fax: +46 18 565705	Order Tel: +1 800 446 4752	jp-1-pointsupport@biotage.com	kr-1-pointsupport@biotage.com	
order@biotage.com Order Fax: +1 704 654 4917				
Support Tel: +46 18 56 59 11	ordermailbox@biotage.com	CHINA	INDIA	
Support Fax: + 46 18 56 57 11	Fax: + 46 18 56 57 11 Support Tel: +1 800 446 4752 itsupport@biotage.com us-1-pointsupport@biotage.com	Tel: +86 21 68162810	Tel: +91 11 45653772	
eu-1-pointsupport@biotage.com		Fax: +86 21 68162829	india@biotage.com	
		cn_order@biotage.com		
		cn-1-pointsupport@biotage.com		

Literature Number: UIUI473

© 2021 Biotage. All rights reserved. No material may be reproduced or published without the written permission of Biotage. Information in this document is subject to change without notice and does not represent any commitment from Biotage. E&OE. A list of all trademarks owned by Biotage AB is available at www.biotage.com/legal. Other product and company names mentioned herein may be trademarks or registered trademarks and/or service marks of their respective owners, and are used only for explanation and to the owners' benefit, without intent to infringe.

